OLYMPUS

List of CDS Related Guidance

Processing Requirements for Thermolabile Flexible Endoscopes



109402

CDS Related Guidance

Processing Requirements for Thermolabile Flexible Endoscopes

Czech Republic

Procedure Room	Decontamination Area				
Precleaning	**Manual Cleaning***	**Manual Disinfection or EWD**	**Drying/Storage/ Sterilization**	**Remarks**	**Source**
		Manually or EWD processing			Vyhláška č. 306/2012 Sb., o podmínkách předcházení vzniku a šíření infekčních onemocnění **(2013)** https://www.zakonyprolidi.cz/cs/2012-306

Legend:
EWD – endoscope washer-disinfector (chemo-thermal mechanical/ automatic cleaning and disinfection)
GDA – glutardialdehyde
PAA – peracetic acid
ClO_2 – chlorine dioxide
LCS – liquid chemical sterilization
EtO – ethylene oxide
LTSF – low-temperature steam formaldehyde
H_2O_2 – hydrogen peroxide
OR – operating room
EHEF – European Hygiene Experts Forum

* Mandatory, even in the case of EWD processing

List of CDS Related Guidance – Czech Republic

infectionprevention.olympus.com

As medical knowledge is constantly growing, technical modifications or changes of the product design, product specifications, accessories and service offerings may be required. This quick reference guide is a mere summary of some important procedure steps and/or product information. It is not meant to replace the instructions for use. Any user of this product must at all times observe all mandatory information for the product, found, in particular, on the labels and the instructions for use.

M00466EN-CZ · 01/22 · OEKG

OLYMPUS

OLYMPUS EUROPA SE & CO. KG
Postbox 10 49 08, 20034 Hamburg, Germany
Wendenstrasse 20, 20097 Hamburg, Germany
Phone: +49 40 23773-0, Fax: +49 40 233765
www.olympus-europa.com